



Test Report

No.: QDHL2010010435MD

Date: OCT.12,2020

Page: 1 of 5

THE TEST RESULTS WERE TRANSFERRED FROM TEST REPORT NO. QDHL2006005323MD.

DATE: JUL.02,2020

Sample Description : DISPOSABLE MEDICAL NITRILE EXAMINATION GLOVES
Style/Item No. : GUG0101A
Lot No. : NOT PROVIDED
Lot Size : NOT PROVIDED
Sample Quantity : 500PCS
Sample Receiving Date : JUN.11,2020
Final Sample Submission Date : JUL.01,2020
Test Performing Date : JUN.11,2020 TO JUL.02,2020

Test Requested

Result

- | | |
|--|------|
| 1. BS EN 455-1:2000 Medical Gloves for Single Use – Part 1: Requirements and Testing for Freedom from Holes (Clause 5.1) | Pass |
| 2. BS EN 455-2:2015 Medical Gloves for Single Use – Part 2: Requirements and Testing for Physical Properties (Clause 4.2, 4.3, 5.2, 5.3) | Pass |
| 3. BS EN 455-3:2015 Medical Gloves for Single Use—Part 3: Requirements and Testing for Biological Evaluation (Clause 4.4, 4.6) | Pass |

Remark: -Unless otherwise stated the results shown in this test report refer only to the sample(s) tested.

SGS-CSTC Standards
Technical Services (Qingdao)
Co., Ltd.

Jessica Gao
Approved Signatory

scan to see the report



QDHL2010010435MD



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.
Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755)83071443, or email: CN.Doccheck@sgs.com

SGS Center, No. 143, Zhuzhou Road, Laoshan District, Qingdao, China 266101
t (86-532) 6899888 f (86-532) 80991955

www.sgs.com
www.sgs.com.cn
e sgs.china@sgs.com

Member of the SGS Group (SGS SA)



Test Report

No.: QDHL2010010435MD

Date: OCT.12,2020

Page: 2 of 5

Test Conducted:

1. BS EN 455-1:2000 Medical gloves for single use – Part 1: Requirements and testing for freedom from holes

Number of test sample	:	200 Pieces
Sample size	:	S
Number of non-conforming gloves	:	0

Clause	Test Items	Result
5	Watertightness test for detection of holes	---
5.1	Referee testing	Pass (See note 1)

Note	:	1	Sample quantity: 200pcs, AQL:1.5, Ac:7, Re:8, Found:0. The sample selecting amount for this clause is deviated to 200 pcs as assessed by SGS.
------	---	---	--

2. BS EN 455-2:2015 Medical gloves for single use – Part 2: Requirements and testing for physical properties

Number of test sample	:	26 Pieces
Type	:	Examination/procedure gloves b)
Size	:	Examination/procedure gloves: S

Clause	Test Items	Result
4	Dimensions	---
4.2	Length	Pass (See result 1)
4.3	Width	Pass (See result 1)
5	Strength	---
5.2	Force at break	Pass (See result 2)
5.3	Force at break after challenge testing	Pass (See result 2)



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755)83071443, or email: CN.Doccheck@sgs.com

SGS Center, No. 143, Zhuzhou Road, Laoshan District, Qingdao, China 266101
t (86-532) 6899888 f (86-532) 80991955

www.sgs.com.cn
e sgs.china@sgs.com

Member of the SGS Group (SGS SA)



Test Report

No.: QDHL2010010435MD

Date: OCT.12,2020

Page: 3 of 5

Result 1: Dimensions

Size	S	
No.	Length (mm)	Width (mm)
1	241	84
2	238	84
3	241	83
4	237	84
5	240	83
6	241	83
7	241	84
8	245	83
9	240	84
10	244	84
11	237	85
12	239	84
13	237	83
Standard requirement	≥240	80±10
Median value	240	84

Result 2: Strength

Size: S			
Force at break (N)			
Before aging		After aging	
No.	/	No.	/
1	8.6	1	7.1
2	9.6	2	9.3
3	10.1	3	8.9
4	8.6	4	6.2
5	9.9	5	8.6
6	8.2	6	8.0
7	8.5	7	6.3
8	8.0	8	6.1
9	9.9	9	9.1
10	7.6	10	6.4
11	8.3	11	8.2
12	5.7	12	4.4
13	8.8	13	7.2
Standard requirement	≥6.0	Standard requirement	≥6.0
Median value	8.6	Median value	7.2



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755)83071443, or email: CN.Doccheck@sgs.com

SGS Center, No. 143, Zhuzhou Road, Lianzhan District, Qingdao, China 266101
t (86-532) 6899888 f (86-532) 80991955

www.sgs.com
sgs.china@sgs.com

Test Report

No.: QDHL2010010435MD

Date: OCT.12,2020

Page: 4 of 5

3. BS EN 455-3:2015 Medical gloves for single use – Part 3: Requirements and testing for biological evaluation

Number of test sample	:	5 Pieces
Sample size	:	S
Finishes of gloves	:	Powdered-free gloves other than surgeon's gloves

Clause	Test Items	Result
4.4	Powder-free gloves	Pass (See note 1)
4.6	Labelling	Pass

Note	:	1	Test according to EN ISO 21171:2006, the average mass of powder per glove is 0.16mg. (Requirement: ≤2mg per powder-free glove)
------	---	---	--

Remark: The declaration of conformity is only based on the actual value of laboratory activity, measurement uncertainty of the results not take into account.

Sample Photo:




Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.
Attention: To check the authenticity of testing /inspection report & certificates, please contact us at telephone: (86-755)83071443, or email: CN.Doccheck@sgs.com

SGS-CSI (Shanghai) Technical Services (Qingdao) Co., Ltd.

SGS Center, No. 143, Zhuzhou Road, Laoshan District, Qingdao, China 266101
t (86-532) 6899888 f (86-532) 80991955

www.sgs.com.cn
e sgs.china@sgs.com



SGS authenticate the photo on original report only

End of Report



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755)83071443, or email: CN.Doccheck@sgs.com

SGS Center, No. 143, Zhuzhou Road, Laoshan District, Qingdao, China 266101
 T (86-532) 6899888 F (86-532) 80991955

www.sgs.com.cn
sgs.china@sgs.com